


austofix Olecranon


3.5mm L&C Plate

Surgical Technique



Contents

Implant Features	5
Plate Features	6
Screw Range	6
• Locking Screw	6
• Spacer	6
• Cancellous Screw	6
• Cortex Screw	6
Surgical Technique	7
• Preparation & Plate Length Selection	7
• Patient Positioning	7
• Incision	7
• Compression Screws	8
• Contouring & Plate Preparation	8
• Pre-Drilling	9
• Position the Plate	9
• Drill Guides	10
• Drilling	11
• Determine Screw Length	12
• Screw Insertion	12
• Cortex/Cancellous Screws	12
• Locking Screws	13
• Spacer	14
Plate Removal	14
Implants	15
• Plates	15
• Screws	16
Instruments	20
Optional Sets	21
• Universal Trauma Set	21
Instrument Trays & Sets	22
Single Use Items	22
MRI Safety	23
Austofix Small Fragment Range	24
Notes	26



Austofix is a leading manufacturer and designer of orthopaedic trauma medical devices with a particular focus on innovation, excellence and patient safety. Austofix has the expertise and experience in developing a new device from concept to a fully Commercialised product with regulatory approval for world-wide distribution.

Throughout its 20+ years Austofix has gathered a team of world-class research and development specialists. Together with orthopaedic surgeons, our specialists identify emerging techniques and innovations in the field of orthopaedic trauma and develop world-class solutions.

Austofix is now one of Australia's key contributors to the world-wide medical technology industry. By focusing on specific market needs we can leverage our staff expertise to develop effective solutions and successfully compete on the world stage.

We understand that accidents don't wait to happen, so we ensure that our equipment and devices are ready when needed. With a dedicated 24 hour, seven day a week customer service and sales team, Austofix products are ready when you are.

With our focus on trauma we understand the specific needs of trauma surgeons. Our product specialists actively support the surgeon by being on call to support procedures and offer advice.

Austofix products and innovations assist the surgeon in performing accurate, efficient and safe procedures that result in better health outcomes for the patient.

The measurement of our success is seen through our excellent clinical results and positive surgeon feedback. We understand the need for efficiency during operations and that this is key in improving patient outcomes. Our products and tools are designed to minimise time spent in theatre. Furthermore, all clinical feedback of our products is promptly addressed to ensure product refinements reflect all surgical concerns.

For further information, updates and contact details visit austofix.com.au and follow us on [LinkedIn](#).

Disclaimer

This document is intended to be read by experienced orthopaedic surgeons familiar with plate fixation

This document is intended as the recommended procedure for using the Small Fragment Plates system. It offers guidance only. Each surgeon should consider the particular needs of the patient and make appropriate adjustments where necessary.

For further advice please contact your local Austofix representative.
© This document is copyright to Austofix and may not be reproduced in whole or part without permission.

austofix Olecranon 3.5mm L&C Plates

The Austofix Olecranon Locking Plates are designed for the fixation of the many fracture patterns found in the Olecranon.

The titanium plates and screws incorporate significant design advantages, facilitating surgical accuracy and efficiency and delivering better patient outcomes.

Austofix understands the importance of proven, high quality medical devices and instruments. The Olecranon L&C Plates adhere to these principles and will provide the surgeon with a comprehensive olecranon fixation solution.

Plate



Screws



Implant Features

Plates

Combi Hole

The Combi Hole allows for a range of plate fixation options. The holes accommodate both Compression and Locking screws.

Slotted Hole - Cortex Screws/Cancellous Screws


Cortex or Cancellous Screws used in the slotted hole for plate-to-bone compression increases stability.

Threaded Hole - Locking Screws

Locking screws link with the threads in the Threaded Hole, keeping the screw at a fixed angle.

Slotted Hole

Threaded Hole



Tapered End

Tapered end assists in submuscular plate insertion and helps to minimise tissue trauma.



Screws


Locking Screw

- Self-Tapping
- Reduced Screw Back-out
- Unicortical or Bicortical Fixation



Cortex(Cortical) Screw

- Dynamic Compression
- Compression



Cancellous Screw

- Dynamic Compression
- Interfragmentary compression (Partially Threaded)
- Compression



Spacer

- Reduce Plate-To-Bone Contact
- Minimises Disruption of Periosteal Blood Supply




Plate Features

Anatomical Fit

- » Pre-contoured for improved anatomical fit on the proximal radius
- » Tapered end assists in submuscular plate insertion and helps to minimise tissue trauma
- » Plate can be contoured with Plate Benders (112100002/3) for a more suitable fit
- » Reconstruction Notches along the shaft of the plate allow for easy reshaping

Proximal Locking


- » Proximal locking holes provide flexibility in Locking Screw fixation
- » Multiple points of fixation for superior angular stability
- » Fixed-angle locking construct providing clinical benefits to patients with osteopenic bone
- » Provides stable fracture fixation while preserving vascular supply to accelerate bone healing
- » Additional fixation of proximal fragments through the use of the removable Proximal Tab

Plate Fixation & Dynamic Compression

- » Combi-Holes along distal shaft of the Plate allow Locking Screws and compression using Cortex and Cancellous screws
- » Gliding Combi-Hole allows adjustment of plate position after preliminary screw insertion
- » Limited-contact distal shaft design
- » Distal Plate shaft has increased thickness for additional strength

Clinical Indications

- » Designed to address simple and complex extra- and intra-articular fractures of the olecranon
- » Can be utilised for repair of the olecranon after osteotomy



Screw Range

Locking Screw




Cancellous Screw



Fully Threaded

Cortex Screw



Spacer



Partially Threaded

Surgical Technique

Preparation & Plate Length Selection

Preoperative radiographic assessment is required to determine plate length. Additionally, the Plate can be contoured to mould to the bone using the supplied Plate Benders (112100002/3).

Patient Positioning


A prone or lateral position is recommended, with the elbow flexed over a support.

A supine position with the forearm placed across the chest is also acceptable.

Incision

The standard site for incision begins posterior from the supracondylar area and extends 4 to 5cm distal to the fracture location.

Note: The incision can be made slightly curved towards the radius to preserve the ulnar nerve.



Reduce the Fracture

Reduce the fracture using the image intensifier, K-Wires* and/or Reduction Forceps (112100011/3). K-Wire bending and manipulation can be achieved using the supplied K-Wire Bender (112200010). Ensure that the reduction instrumentation will not interfere with plate placement.

***Note:** 1.4mm (511415) and 2.0mm (522015) K-Wires are available for fracture reduction.

Suture Reduction (Optional)

Sutures can be used through the K-Wire holes to reposition tuberosity fragments, further stabilising the construct and providing provisional fixation. The sutures are then attached to the plate.

Compression Screws

Cannulated Compression Screws (Ø3.0 - Ø7.3mm) are available for interfragmentary compression and fracture fixation.


Please refer to the Austofix Cannulated Compression Screws Surgical Technique.

Note: The Cannulated Compression Screw instrument set (SET-INS-CAN) is required for this approach.

Headless Cannulated Screws are also available for fixation of intra-articular and extra-articular fractures, and non-unions of small bones.

Please refer to the Austofix Headless Cannulated Screws Surgical Technique.

Note: The Headless Cannulated Screw instrument set (SET-INS-HLCANN) is required for this approach.




Contouring & Plate Preparation

If contouring is required, make sure to place the Plate Benders (112100002/3) on two consecutive holes to avoid distortion of the Threaded Holes. Bending Templates (Large: 112100008 & Small: 112100009) are included if desired.

The plate can be bent at a maximum of 4° at each Reconstruction Notch.


Warning: Do NOT bend the plate beyond what is required to contour with the bone. Reverse bending, over bending, using the incorrect instrumentation for bending or bending at the level of the holes may lead to plate failure.



Proximal Tab

The Proximal Tab of the L&C Olecranon Locking Plate may be utilised for additional fixation of proximal fragments. The Tab can be bent for optimal Screw positioning. The Proximal Tab may also be removed if necessary using appropriate plate cutting instrumentation.

Note: If the Proximal Tab is bent, ensure proximal Locking Screws will not intersect.



Pre-Drilling

Determine whether Cortex, Cancellous or Locking Screws will be used. A combination of Screws may also be used.

Use the table to determine which combination of Drill/Drill Sleeve is required for the desired Screw.

Note: If a combination of Screws is used, a Cortex Screw should be inserted first to pull the plate to the bone.


Note: If a Locking Screw is used first, care should be taken to ensure that the plate is securely held to the bone to avoid spinning of the plate about the bone as the Locking Screw is tightened to the plate.

Screw	3.5mm Cortex	4.0mm Cancellous	3.5mm Locking
Drill	Ø2.5mm Drill (112100016)	Ø2.5mm Drill (112100016)	Ø2.8mm Drill (112200004)
Drill Sleeve/ Guide	2.5mm Drill Guide (112100020)	2.5mm Drill Guide (112100020)	2.8mm Threaded Drill Sleeve (112200002)
Driver	SW2.5 Hex Screwdriver (112100022/ 112100017)	SW2.5 Hex Screwdriver (112100022/ 112100017)	T15 Star Screwdriver (112200009/ 112200003)
Torque Limiter	-	-	1.5Nm Torque Limiter (112200001)

Position the Plate

Place the selected plate on the fractured bone and in a suitable position. If axial dynamic compression will be used ensure the middle of the plate is above the line of the fracture. The plate can be temporarily held in place using plate holding forceps. 2.0mm K-Wires (522015) can be used to assist with determining the optimal position of the plate.

Once the plate has been positioned appropriately, a preliminary Cortex Screw should be inserted through the elongated slotted hole of the Gliding Combi-Hole.




Drill Guides

The Small Frag Instrument Set contains four different drill guides. Please follow the guide below to ensure the drill guides are used appropriately.

Drill Guide, LC-L&C 2.5/3.5mm (112100004)


The Limited Contact Locking & Compression (LC-L&C) Drill Guide is only used with plates with limited contact features and bidirectional **Dynamic Compression Holes**. The sleeve tips are designed to seat precisely within the hole to allow accurate neutral or compression drilling.

Warning: The LC-L&C Drill Guide 2.5/3.5 (112100004) is **not** suitable for use with the 3.5mm L&C Olecranon Plate listed in this Surgical Technique as no **bidirectional Dynamic Compression Holes** are present. Please use the Universal Drill Guide 3.5/2.5mm (112100020) to drill through **Combi-Holes**.




Drill Sleeve, Double 2.5/3.5mm (112100005)

The Double Drill Sleeve is a standard drill sleeve used predominantly for inserting Screws outside of the plate to capture fragments. The larger diameter drill sleeve can additionally be used as a tap sleeve.



Drill Guide, Universal 3.5/2.5mm (112100020)

The Universal Drill Guide can be used for insertion of Cortex and Cancellous Screws (**not** Locking Screws) through all holes including **Combi-Holes** found in the Small Fragment plating range. The spring-loaded tip allows for greater control over Screw angle within the plate hole. The drill guide also contains a larger diameter tap sleeve if tapping is required. Please refer to the Drilling section (page 11) for more information on using this drill guide.



Drill Sleeve, Threaded 2.8mm (112200002)

The Threaded Drill Sleeve is used within the threaded portion of the **Combi-Holes** and the threaded **Locking Holes** for perpendicular insertion of Locking Screws for angular stability.



Drilling

Using the Universal Drill Guide 3.5/2.5mm

Neutral Screw Position

Advance the spring-loaded Universal Drill Guide 3.5/2.5mm (112100020) through the Dynamic Compression slot of the plate. Press the Drill Guide against the bone, allowing the inner sleeve to retract. This will guide the rounded end of the outer sleeve to the Neutral drilling position.

Dynamic Screw Position


Place the inner sleeve of the Universal Drill Guide 3.5/2.5mm (112100020) against the edge of the Dynamic Compression slot of the plate. Without exerting pressure on the Drill Guide, the inner sleeve will remain in the Dynamic position. Dynamic Compression will occur once the Cortex/Cancellous Screws are tightened (page 12).

Locking Sleeve & Locking Screw Position

Insert the 2.8mm Threaded Drill Sleeve (112200002) into threaded hole. Carefully drill the Locking Screw hole using the 2.8mm Drill (112200004).


Note: The 2.8mm Threaded Drill Sleeve and the Ø2.8mm Drill have a **blue** laser marking for easy identification.

Warning: Take care while drilling as interference can occur between Screws in the proximal portion of the plate. If crossing over between drilled holes occurs, cease drilling and use a Screw of the appropriate length.



Determine Screw Length


Locking Screw length can be determined by reading the depth measurement directly from the calibrated Ø2.8mm Drill (112200004).



The 2.5/4.0mm Depth Gauge (112100001) is required to determine Cortex and Cancellous Screw length. The Depth Gauge can be inserted directly into the hole in the bone. Measurements marked on the Depth Gauge are used to determine the Screw length. The Depth Gauge can also be used as an alternative to determining Locking Screw length.

Insert the hook of the Depth Gauge to engage the dorsal Cortex of the bone.

Read the measurements from the barrel of the Depth Gauge. If the measurement is between graduations choose the smaller Screw length.



Screw Insertion

Select the appropriate Screw with the assembled Driver Tip and Handle. Self holding driver tips are used to secure the screw to the appropriate driver.

Note: Use the table on page 9 to determine which Driver and Handle to use.

Note: If a combination of Cortex and Locking Screws are to be used then the Cortex Screw should be inserted first to generate interfragmentary compression.



Cortex/Cancellous Screws

Neutral Screw Position


Use the Hex Screwdriver (112100022) or Power Screwdriver (112100017) to insert Cortex/Cancellous Screw to appropriate depth.

Alternatively, a Power Screwdriver (112100017) can be used.

Dynamic Screw Position

Use the Hex Screwdriver (112100022) or Power Screwdriver (112100017) to insert Cortex/Cancellous Screw to appropriate depth. The Screwdriver's Holding Sleeve can be used to assist in Screw insertion. Ensure Dynamic Compression has been achieved.

Warning: If using power to insert Cortex/Cancellous Screws, complete Screw insertion by hand to avoid compromising the integrity of the plate and screw construct, and to avoid damage to patient soft tissue.



Neutral Screw Position

Dynamic Compression Position


Partially Threaded vs Fully Threaded Cancellous Screws

Partially Threaded

- Used to compress the plate to the bone and can be used to provide interfragmentary compression

Fully Threaded

- Used to compress the plate to the bone




Locking Screws

Before the insertion of the Locking Screw ensure the fracture is reduced as this cannot be done after the screw has been inserted.

Ensure the desired Screw is concentric to the plate's threaded hole. Insert the Screw and tighten with the 1.5Nm Torque Limiter Screwdriver (112200001). Screw it down until the threaded screw head engages and is secure to the plate.

Warning: If using power to insert Locking Screws, always use a torque limiting attachment. This reduces the risk of the threads stripping from the head of the screw. The Star Screwdriver Shaft (112200003) can be used for insertion using power. The Locking Screws should always be tightened to the final position by hand.




Depending on the clinical requirements, no compression or dynamic compression will be generated. The number of Screws and Screw insertion configuration is to be determined by the surgeon.

Warning: If Locking Screws in the most proximal portion of the plate exceed 30mm, interference with Screws in the distal shaft can occur.

Spacer

If minimal contact between the plate and bone is desired, a 3.5mm Spacer can be inserted using the SW2.5 Hex Driver (112100022). The Spacer ensures a distance of 2mm between the plate and bone when the Screws are later inserted. The Spacer can be removed after setting the Locking Screws.




Spacer



Plate Removal

To remove the Plate, unlock and partially unscrew all the Screws first using the appropriate Hex (112100022) or Star (112200009) Screwdriver for Cortex/Cancellous Screws and Locking Screws respectively. Continue to remove the Screws from the bone. This method prevents the simultaneous rotation of the plate when unlocking the final Locking Screw.

Note: The Easyout (112200005) can be used with the T-Handle with Quick Coupling (112100024) or a Power Drill if there are difficulties in the removal of Screws.




Implants

Plates

L&C Olecranon Locking Plate		
Product Code	Number of Holes	L/R
2055-00-02086L	2	L
2055-00-04112L	4	L
2055-00-06138L	6	L
2055-00-08163L	8	L
2055-00-10190L	10	L
2055-00-12216L	12	L
2055-00-02086R	2	R
2055-00-04112R	4	R
2055-00-06138R	6	R
2055-00-08163R	8	R
2055-00-10190R	10	R
2055-00-12216R	12	R


Compatible Screw: 3.5mm Cortex Screw, 4.0mm Cancellous Screw, 3.5mm Locking Screw, 3.5mm Spacer

Compatible K-wire: 2.0mm




Screws

Locking Screw - Self-tapping	
Length	3.5mm
10	1061-00-35010
12	1061-00-35012
14	1061-00-35014
16	1061-00-35016
18	1061-00-35018
20	1061-00-35020
22	1061-00-35022
24	1061-00-35024
26	1061-00-35026
28	1061-00-35028
30	1061-00-35030
32	1061-00-35032
34	1061-00-35034
35	1061-00-35035
36	1061-00-35036
38	1061-00-35038
40	1061-00-35040
42	1061-00-35042
44	1061-00-35044
45	1061-00-35045
46	1061-00-35046
48	1061-00-35048
50	1061-00-35050
55	1061-00-35055
60	1061-00-35060
65	1061-00-35065
70	1061-00-35070
75	1061-00-35075
80	1061-00-35080
85	1061-00-35085
90	1061-00-35090



Cortex Screw - Self-tapping	
Length	3.5mm
12	1145-00-35012
14	1145-00-35014
16	1145-00-35016
18	1145-00-35018
20	1145-00-35020
22	1145-00-35022
24	1145-00-35024
26	1145-00-35026
28	1145-00-35028
30	1145-00-35030
32	1145-00-35032
34	1145-00-35034
36	1145-00-35036
38	1145-00-35038
40	1145-00-35040
42	1145-00-35042
44	1145-00-35044
46	1145-00-35046
48	1145-00-35048
50	1145-00-35050



Cancellous Screw - Partially-threaded


Length	4.0mm
20	1146-00-40020
22	1146-00-40022
24	1146-00-40024
26	1146-00-40026
28	1146-00-40028
30	1146-00-40030
32	1146-00-40032
34	1146-00-40034
36	1146-00-40036
38	1146-00-40038
40	1146-00-40040
42	1146-00-40042
44	1146-00-40044
46	1146-00-40046
48	1146-00-40048
50	1146-00-40050
52	1146-00-40052
54	1146-00-40054
56	1146-00-40056
58	1146-00-40058
60	1146-00-40060




Cancellous Screw - Fully-threaded	
Length	4.0mm
12	1147-00-40012
14	1147-00-40014
16	1147-00-40016
18	1147-00-40018
20	1147-00-40020
22	1147-00-40022
24	1147-00-40024
26	1147-00-40026
28	1147-00-40028
30	1147-00-40030
32	1147-00-40032
34	1147-00-40034
36	1147-00-40036
38	1147-00-40038
40	1147-00-40040
42	1147-00-40042
44	1147-00-40044
46	1147-00-40046
48	1147-00-40048
50	1147-00-40050
52	1147-00-40052
54	1147-00-40054
56	1147-00-40056
58	1147-00-40058
60	1147-00-40060



Spacer	
Length	3.5mm
2	1060-00-35002



Instruments




Instruments

#	Code	Description	Qty	#	Code	Description	Qty
1	112200001	Torque Limiter 1.5Nm	1	12	112200003	Screwdriver Shaft, 120mm (Star)	1
2	112100022	Screwdriver (Hex) with Holding Sleeve	1	13	112100015	Drill Bit 3.5mm	1
3	112200009	Screwdriver, 200mm (Star)	1	14	112100016	Drill Bit 2.5mm	2
4	112100001	Depth Gauge 2.5/4.0mm	1	15	112100014	Countersink, 100mm	1
5a.	112100002	Plate Bender (Left), 190mm	1	16	112100017	Screwdriver Shaft 100mm (Hex)	1
5b.	112100003	Plate Bender (Right), 190mm	1	17	112100018	Tap for Cortex Screw 3.5mm	1
6	112100004	Drill Guide, LC-L&C 2.5/3.5mm	1	18	112100019	Tap for Cancellous Bone Screw 4.0mm	1
7	112100005	Drill Sleeve, Double 2.5/3.5mm	1	19a.	112100008	Bending Template (Large) 10 x 118mm	1
8	112100020	Drill Guide, Universal 3.5/2.5mm	1	19b.	112100009	Bending Template (Small) 10 x 92mm	1
9	112200005	Easyout, 80mm	1	20	112200004	Drill Bit 2.8mm	2
10	112200002	Drill Sleeve, Threaded, 2.8mm (for 3.5)	2	21	112200010	K-Wire Bender	1
11	112100024	T-Handle with Quick Coupling, 90mm	1	22	112127000	Small Frag Screw Tray	1

Optional Sets

Universal Trauma Set



Instruments			
#	Code	Description	Qty
1	113100017	Hohmann Retractor (Large) 43.5 x 267mm	2
2	113100019	Self-Centering Bone Holding Forceps (Speed Lock) 266mm	1
3	113100021	Reduction Forceps (Serrated Jaws) 220mm	2
4	113100022	Reduction Forceps (Point) 207mm	1
5.	113100018	Hohmann Retractor (Small) 16 x 267mm	1
6	112100010	Self-Centering Bone Holding Forceps (Speed Lock) 191.8mm	2
7	112200012	Self-Centering Bone Holding Forceps (Compression)	1
8	112100011	Reduction Forceps (Serrated Jaws) 158mm	1
9	112100013	Reduction Forceps (Points) 182mm	1
10	112100006	Hohmann Retractor (Large) 15.5 x 159mm	2
11	112100007	Hohmann Retractor (Small) 10.5 x 170mm	2
12	113100016	Periosteal Elevator (Large) 191mm	1
13	112100012	Periosteal Elevator (Small) 190mm	1

Instrument Trays & Sets

Instrument Tray		
Code	Description	Qty
112128000	Small Frag Instrument Tray PPSU (Empty)	1
113122000	Universal Trauma Instrument Tray (Empty)	1

Instrument Set		
Code	Description	Qty
SET-INS-SML	Full Small Frag Instrument Set	-
SET-INS-UTRA	Full Universal Trauma Instrument Set	-

Single Use Items

Recommended K-Wires		
Code	Description	Qty
522015	2.0 x 150mm K-Wire	2
511415	1.4 x 150mm K-Wire	2



Optional K-Wires		
Code	Description	Qty
113210001	2.5 x 280mm K-Wire	2
611.112	1.1 x 120mm K-Wire	2
081.010	0.8 x 100mm K-Wire	2



Drill		
Code	Description	Qty
112100016	Drill Bit 2.5mm (for 3.5mm Cortex & 4.0mm Cancellous Screws)	1
112200004	Drill Bit 2.8mm (for 3.5mm Locking Screws)	1
112100015	Drill Bit 3.5mm	1



MRI Safety

Austofix has not evaluated its devices for safety and compatibility in a Magnetic Resonance (MR) environment. However, the materials used in their manufacture are known to have minimal ferromagnetism, with minimal risk to patients in strong magnetic fields.

Austofix has performed a review of published, peer-reviewed data, which confirms that only minor rises in MRI-related heating are observed from devices manufactured from the same titanium and stainless-steel materials. Trauma devices are considered unlikely to produce injury to patients, including in the worst-case 3.0T systems.

The devices and materials observed in the literature experience forces too weak to cause significant displacement; the risk being further mitigated by their implantation in bone. Risks of imaging artifacts are known to MRI operators, and can be reduced by choosing appropriate pulse sequences and optimizing scanning parameters by using a large bandwidth, small field-of-view and appropriate echo train length.

Average temperature changes have been observed in studies at 0.48°C in titanium and 0.74°C in stainless-steel. Rises in temperature in clinical situations may depend on individual patient factors. It should be recommended that patients be thoroughly monitored when undergoing MR scanning, and that impaired patient thermoregulation be considered a contraindication for MRI procedures.

Sources:

Chen CA, Chen W, Goodman SB, et al. New MR Imaging Methods for Metallic Implants in the Knee: Artifact Correction and Clinical Impact. 2011, 1121-1127.

Gill A, Shellock FG. Assessment of MRI issues at 3-Tesla for metallic surgical implants: findings applied to 61 additional skin closure staples and vessel ligation clips. J Cardiovasc Magn Reson. 2012, 14(1):3.


Shellock FG. Biomedical Implants and Devices: Assessment of Magnetic Field Interactions With a 3.0-Tesla MR System. 2002, 721-732.

Zou Y, Chu B, Wang C, Hu Z. Evaluation of MR issues for the latest standard brands of orthopedic metal implants, Plates and screws. Eur J Radiol. 2015, 84(3):450-457.

Austofix Small Fragment Range

The Austofix Small Fragment Instrument Set and the Austofix Universal Trauma Set are compatible with the entire Austofix Small Fragment range of plates listed below.

For more information on the usage and technique of these plates or for product codes, see the relevant plate-specific Surgical Technique.



L&C Proximal Humeral Locking Plate

L&C Hook Locking Plate

L&C Superior Anterior Clavicle Locking Plate

L&C Superior Anterior Clavicle Lateral Extension Locking Plate

Clavicle Reconstruction Locking Plate

L&C Distal Lateral Dorsal Humeral Locking Plate (Buttress)

L&C Distal Medial Humeral Locking Plate

L&C Distal Lateral Dorsal Humeral Locking Plate


L&C Distal Humeral Extra-Articular Locking Plate

L&C Distal Humeral Middle Metaphyseal Locking Plate

L&C Olecranon Locking Plate

Proximal Humeral Greater Tubercle Locking Plate

Continued



L&C Distal Medial Tibial Locking Plate

L&C Clavicle Hook Locking Plate

L&C Proximal Posterior Tibial Locking Plate

L&C Distal Posterior Lateral Fibula Locking Plate

L&C Distal Lateral Fibula Locking Plate

L&C Proximal Medial Tibial Locking T-Plate

1/3 Tubular Locking Plate


3.5mm L&C Metaphysis Locking Plate

Calcaneal Locking Plate I

Calcaneal Locking Plate II (53mm)

Calcaneal Locking Plate II (60mm)

Calcaneal Locking Plate II (68mm)



3.5mm L&C Reconstruction Locking Plate (Straight)

3.5mm LC-L&C Locking Plate (Narrow)

3.5mm L&C Locking T-Plate Right-Angle (Head 4 Hole)

3.5mm L&C Locking T-Plate Oblique-Angle (Head 3 Hole)

Clavicle Anterior Reconstruction Locking Plate

3.5mm L&C T-Plate Right-Angle (Head 3 Hole)

L&C Proximal Radius Arch Rising Locking Plate

L&C Proximal Radius Arch Cupped Locking Plate

L&C Anterolateral Distal Tibial Locking Plate

3.5mm L&C Distal Medial Tibial Locking Plate

L&C Proximal Lateral Tibial Locking Plate I

L&C Proximal Lateral Tibial Locking Plate



Notes

Notes



austofix

Legal Manufacturer

Australian Orthopaedic Fixations Pty Ltd
18 Kinkaid Avenue, North Plympton SA 5037, AUSTRALIA

Telephone: 1300 727 075
Fax: 1300 727 380
Email: sales@austofix.com.au

DISTRIBUTOR STICKER



Advena Ltd. Tower Business
Centre, 2nd Flr., Tower Street,
Swatara, BKR 4013 Malta



austofix.com.au



700104 Olecranon ST Rev 2